



**Georgia Department of Agriculture
Fuel & Measures Division
Capitol Square, Atlanta, Georgia**

Date 6-3-84
Time Start 07050
Time End 01110

3249

Tommy Irvin
Commissioner

VEHICLES-DORMANT-POULTRY TEST REPORT

Telephone: 404-656-7057
Fax: 404-656-9648

TYPE OF INSPECTION:

- Change New Company New Device
 Regular Reinspection Request

Owner No.: <u>0210012</u>	Company Name: <u>Macon Iron & Paper Stock</u>	Street Address: <u>160 Edgewood Ave.</u>
City: <u>Macon</u>	County: <u>Bibb</u>	Zip: <u>31206</u>
Device Number: <u>02095</u>	Make of Scale: <u>GSE-550</u>	Capacity: <u>60 Ton</u>
Change <input type="checkbox"/> Delete <input type="checkbox"/>	Last Test Date: <u>5-8-83</u>	Bal. Condition: <u>0-</u>
		Kind of Beam or Dial: <u>Electronic</u>
		Min. Grad.: <u>10lb</u>
		Printer: <u>YES</u>
		Pit Depth: <u>—</u>
		Platform Size: <u>10'x60'</u>
		Witness Signature: _____

TEST RESULTS - INCREASING LOAD TEST

SR AT ZERO LOAD _____ LBS.					SR AT MAXIMUM LOAD _____ LBS.				
TEST WEIGHTS					TEST WEIGHTS				
Position	Balance	POUNDS	SCALE INDICATION	ERROR	Position	Balance	POUNDS	SCALE INDICATION	ERROR
<u>Sec3</u>		<u>3000</u>	<u>2980</u>	<u>-20</u>	<u>Sec3</u>		<u>21000</u>	<u>20960</u>	<u>-40</u>
		<u>5000</u>	<u>4980</u>	<u>-20</u>	<u>Sec2</u>		<u>21000</u>	<u>20980</u>	<u>-20</u>
		<u>7000</u>	<u>6980</u>	<u>-20</u>	<u>Sec1</u>		<u>21000</u>	<u>20980</u>	<u>-20</u>
		<u>9000</u>	<u>8980</u>	<u>-20</u>	<u>RTZ</u>	<u>0</u>	<u>0</u>	<u>0</u>	<u>0</u>
		<u>12000</u>	<u>11980</u>	<u>-20</u>	<u>RTZ</u>	<u>0</u>	<u>0</u>	<u>0</u>	<u>0</u>
<u>Sec2</u>		<u>12000</u>	<u>11980</u>	<u>-20</u>					
<u>Sec1</u>		<u>12000</u>	<u>11980</u>	<u>-20</u>					
<u>Sec3</u>		<u>15000</u>	<u>14980</u>	<u>-20</u>					
		<u>18000</u>	<u>17980</u>	<u>-20</u>					

STRAIN LOAD TEST

	Section 1	Section 2	Section 3	Section 4	Section 5	Section 6
INDICATE WEIGHT OF EMPTY TRUCK	<u>26520</u>	<u>26520</u>				
AMOUNT OF TEST WEIGHTS ADDED	<u>21000</u>	<u>21000</u>				
SCALE INDICATION, TRUCK PLUS WEIGHTS	<u>47500</u>	<u>47500</u>				
ERROR TEST WEIGHTS ADDED	<u>-20</u>	<u>-20</u>				

ACTION: (1) APPROVED (2) REJECTED (3) CONDEMNED (4) TIED DOWN (5) OUT OF USE (7) NO TEST

DEVICE USE: 13 DEVICE TYPE: 13

INSPECTION INTERVAL: 01

REMARKS: _____

Jed Sweet Earl Oute 23
Inspector Signature Number

[Signature]
Receipt of Report Acknowledge (Owner/Agent)



GEORGIA DEPARTMENT OF AGRICULRE

FUEL AND MEASURES

19 MLK RM 321 ATLANTA, GEORGIA 30334-4201
PHONE 404-656-3605 FAX 404-656-9648

Device Detail

6006U

Insp Date: 4/5/2007 Business ID: 0210124
Business: MACON IRON & PAPER STOCK
181 WELDWOOD LANE
MACON, GA 01208

Inspection: 83000121
Store #:
Phone: 478-745-1143
Inspector: 48A MIKE TANNER
Reason: Routine

Make: GSE
Model: 550
Serial: 020497
Capacity: 100000
Class: III L

Type: Heavy Capacity Scale
Subtype: Procedure B 3 section
Insp Type: Maintenance
Loc: BESIDE SCALE IN OFFICE
Units: lb
Device Use: Metal-13

Results: Approved
Seal #:
Other:
d: 20
Device Type: Vehicle-13

Type	Name	Display	Inc/Base	Actual	Err	Tol	Res	Notes
Test	Zero	0	0	0	0	20	Pass	
Test	Increasing Load	3020	0	3000	20	20	Pass	
Test	Increasing Load	5020	0	5000	20	20	Pass	
Test	Increasing Load	7020	0	7000	20	20	Pass	
Test	Increasing Load	9020	0	9000	20	20	Pass	
Test	Increasing Load	12020	0	12000	20	40	Pass	
Test	Shift	12000	0	12000	0	40	Pass	
Test	Shift	12000	0	12000	0	40	Pass	
Test	Increasing Load	15020	0	15000	20	40	Pass	
Test	Increasing Load	18020	0	18000	20	40	Pass	
Test	Increasing Load	21020	0	21000	20	60	Pass	
Test	Shift	21020	0	21000	20	60	Pass	
Test	Shift	21000	0	21000	0	60	Pass	
Test	Zero	0	0	0	0	20	Pass	
Test	Strain Load	21000	0	21000	0	60	Pass	
Test	Decreasing Load	21020	0	21000	20	60	Pass	
Test	Zero	0	0	0	0	20	Pass	
Attrib	Motion Detector						Pass	
Attrib	Approches						Pass	
Attrib	AZSM						Pass	
Attrib	Capacity						Pass	
Attrib	Clearance						Pass	
Attrib	Customer Readability						Pass	
Attrib	Marking						Pass	
Attrib	Suitability of Equipment						Pass	

Notes:

Official

Acknowledged Receipt: Michael K Coates



GEORGIA DEPARTMENT OF AGRICULTURE

FUEL AND MEASURES

19 MLK RM 321 ATLANTA, GEORGIA 30334-4201
 PHONE 404-656-3605 FAX 404-656-9648


Device Detail

map date: 4/30/2007 business id: 0210092 inspection: 65000120
 Business: MACON IRON & PAPER STOCK Store #:
 950 LOWER POPULAR ROAD Phone: 487-743-6773


MACON, GA 31208 Reason: Routine

Make: RICE LAKE Type: Meter Capacity Scale Results: Approved
 Model: 420 PLUS-2A Subtype: Procedure B 4 section Seal #:
 Serial: 4388408055 Insp Type: Maintenance Other:
 Loc: OFFICE
 Capacity: 120000 Units: lb G: 10
 Class: III Device Use: Metal-13 Device Type: Vehicle-13

Type	Name	Display	Inc/Base	Actual	Err	LOI	Res	Notes
Test	Zero	0	0	0	0	10	Pass	
Test	Increasing Load	3010	0	3000	10	10	Pass	
Test	Increasing Load	6010	0	6000	10	10	Pass	
Test	Increasing Load	7010	0	7000	10	20	Pass	
Test	Increasing Load	9010	0	9000	10	20	Pass	
Test	Increasing Load	12010	0	12000	10	20	Pass	
Test	Shift	12010	0	12000	10	30	Pass	
Test	Shift	12010	0	12000	10	30	Pass	
Test	Shift	12020	0	12000	20	30	Pass	
Test	Increasing Load	15020	0	15000	20	30	Pass	
Test	Increasing Load	18020	0	18000	20	40	Pass	
Test	Increasing Load	21030	0	21000	30	50	Pass	
Test	Increasing Load	21030	0	21000	30	50	Pass	
Test	Increasing Load	21020	0	21000	20	50	Pass	
Test	Zero	0	0	0	0	10	Pass	
Test	Strain Load	21020	0	21000	20	50	Pass	
Test	Decreasing Load	21020	0	21000	20	50	Pass	
Test	Zero	0	0	0	0	10	Pass	
Test	Increasing Load	0	0	0	0	10	Pass	
Attrib	Motion Detector						Pass	
Attrib	Approches						Pass	
Attrib	Capacity						Pass	
Attrib	Clearance						Pass	
Attrib	Customer readability						Pass	
Attrib	Marking						Pass	



 Official



 Acknowledged Receipt: Jeffery Nebel